


Precision alcoholmeters class I and II

Grey colour code = suitable for official testing

Additional order no.:

-/01 officially calibrated without certificate
-/02 officially calibrated with certificate
-/03 with works certificate
-/04 with DKD calibration certificate

Our alcoholmeters are accredited by Physikalisch-Technische Bundesanstalt (PTB). For larger pictures of the accreditation certificates see page 46 in the technical information chapter.



Precision alcoholmeters

Class I – EG – OIML

Ref. temperature (Tp.): 20 °C

Measuring span: 10% vol

Scale units: 0.1% vol

Total length (L): 480 mm

Range % mas	Ref.- No.
0-10	3302085
10-20	3302086
20-30	3302087
30-40	3302088
35-45	3302089
40-50	3302090
50-60	3302091
60-70	3302092
70-80	3302093
80-90	3302094
90-100	3302095
95-103	3302096

Precision alcoholmeters

Class II – EG – OIML

Ref. temperature (Tp.): 20 °C

Measuring span: 10% vol

Scale units: 0.1% vol

Total length (L): 350 mm

Range % mas	Ref.- No.
0-10	3302061
10-20	3302062
20-30	3302063
30-40	3302064
35-45	3302065
40-50	3302066
50-60	3302067
60-70	3302068
70-80	3302069
80-90	3302070
90-100	3302071
95-103	3302072



Precision alcoholmeters

Class II – EG – OIML

Ref. temperature (Tp.): 20 °C

Measuring span: 10% vol

Scale units: 0.1% vol

With thermometer

(+5 to +25 °C in 0.2 °C)

Total length (L): 390 mm

Range % mas	Ref.- No.
0-10	3302073
10-20	3302074
20-30	3302075
30-40	3302076
35-45	3302077
40-50	3302078
50-60	3302079
60-70	3302080
70-80	3302081
80-90	3302082
90-100	3302083
95-103	3302084

